

Book Chapter

A Microfluidic Spheroid Culture Device with a Concentration Gradient Generator for High-Throughput Screening of Drug Efficacy

Wanyoung Lim¹ and Sungsu Park^{1,2,*}

¹Department of Biomedical Engineering, Sungkyunkwan University, Korea

²School of Mechanical Engineering, Sungkyunkwan University, Korea

***Corresponding Author:** Sungsu Park, Department of Biomedical Engineering, Sungkyunkwan University, Suwon 16419, Korea

Published **October 04, 2021**

This Book Chapter is a republication of an article published by Wanyoung Lim and Sungsu Park at *Molecules* in December 2018. (Lim, W.; Park, S. A Microfluidic Spheroid Culture Device with a Concentration Gradient Generator for High-Throughput Screening of Drug Efficacy. *Molecules* 2018, 23, 3355. <https://doi.org/10.3390/molecules23123355>)

How to cite this book chapter: Wanyoung Lim, Sungsu Park. A Microfluidic Spheroid Culture Device with a Concentration Gradient Generator for High-Throughput Screening of Drug Efficacy. In: Letizia Giampietro, editor. *Prime Archives in Molecular Sciences: 2nd Edition*. Hyderabad, India: Vide Leaf. 2021.

© The Author(s) 2021. This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits

unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Author Contributions: W.L. and S.P.; conceived, designed the study, W.L.; performed and analysed the experiments, W.L. and S.P.; wrote the paper, S.P.; supervised the whole research.

Funding: Technology Innovation Program (or Industrial Strategic Technology Development Program-Development of disease models based on 3D microenvironmental platform. mimicking multiple organs and evaluation of drug efficacy) (20008413) funded By the Ministry of Trade, Industry & Energy (MOTIE, Korea)..

Conflicts of Interest: The authors declare no conflict of interest.

Abstract

Three-dimensional (3D) cell culture is considered more clinically relevant in mimicking the structural and physiological conditions of tumors in vivo compared to two-dimensional cell cultures. In recent years, high-throughput screening (HTS) in 3D cell arrays has been extensively used for drug discovery because of its usability and applicability. Herein, we developed a microfluidic spheroid culture device (μ FSCD) with a concentration gradient generator (CGG) that enabled cells to form spheroids and grow in the presence of cancer drug gradients. The device is composed of concave microwells with several serpentine micro-channels which generate a concentration gradient. Once the colon cancer cells (HCT116) formed a single spheroid (approximately 120 μ m in diameter) in each microwell, spheroids were perfused in the presence of the cancer drug gradient irinotecan for three days. The number of spheroids, roundness, and cell viability, were inversely proportional to the drug concentration. These results suggest that the μ FSCD with a CGG has the potential to become an HTS platform for screening the efficacy of cancer drugs.

Keywords

Spheroids; Concentration Gradient Generator; Drug Screening; High-Throughput

Introduction

In general, two-dimensional (2D) cell culture models have been used for evaluating toxicity or effectiveness of drug candidates [1]. However, it has been shown that (a) 2D cell culture models are relatively poor in predicting drug responses, and (b) there are differences in the functional and phenotypic features between 2D and 3D cell cultures [2,3]. 3D cell culture models represent the *in vivo* microenvironments more accurately, and their predictability of drug effectiveness is better [4–6]. Recently, high-throughput screening (HTS) techniques of drugs have incorporated 3D cell cultures, and have rapidly progressed in the testing and selection of these drugs [7–9]. Accordingly, 3D cell arrays are extensively used nowadays for drug screening applications. From these, scaffold-free, 3D cell arrays—such as low-adhesion plates, micro-patterned plates, and hanging drop micro-plates—are the most commonly used at the present time [10–13]. The self-aggregation of cells is the key point of these methods. Moreover, these 3D cell arrays would include scaffolds, such as hydrogels and meshes, to mimic cell-to-extracellular matrix interactions and tissue-specific properties [14,15]. These micro-plate-based 3D cell arrays need complicated equipment such as robot arm, detectors and software for handling solution and data processing [16]. Furthermore, they are associated with several advantages, such as their reproducibility, simplicity of use for handling cultures, and abilities to treat and routinely analyse multi-cellular spheroids. However, this robotic equipment still suffers from several technical issues, such as high costs, poor handling of small volumes of liquid, and cell culture contamination [17]. The cost of devices and reagents for drug screening is also high because of high-volume consumptions. Therefore, new HTS systems combined with microfluidics, which require low sample and liquid volumes, are affordable, can easily handle small liquid

volumes, enable serial processing and analysis, and are in the immediate needs for the drug development industry.

Microfluidic spheroid formation platforms have been applied to HTS for long-term perfusion cell cultures and have maintained high-cell viability. In the past, numerous microfluidic systems have been developed for formation of spheroids using microwells or U-shaped microstructures in the device. Microwell-based microfluidic platforms have been utilised more than other methods owing to their simplicity and easy operation [18–20]. Liu et al. designed a microfluidic device with U-shaped microstructures for spheroid formation and HTS [21]. These platforms were often combined with a concentration gradient generator (CGG) as a mixing channel [22–24]. Such a channel can be controlled for precise flow control. Recently, Fan et al. reported a high-throughput drug screening brain cancer chip composed of a photo-polymerised hydrogel to form multiple cancer spheroids [25]. They demonstrated that the culture array in association with a gradient generator was capable of forming spheroids, and for widespread parallel testing of drug responses. However, their microfluidic chip is difficult to be commercialised because of the short storage time of the hydrogel. In addition, because cells are injected through inlets, it is difficult for cells to go into the microwells through sub-channels, and their losses are thus high.

In this study, we developed a microfluidic spheroid culture device (μ FSCD) with a CGG for evaluating the efficacy of cancer drugs to spheroids. In the device made of polydimethyl siloxane (PDMS), colon cancer cells (HCT116) were first deposited in concave microwells and later adhered to each other because they were not able to attach to the surface that was coated with bovine serum albumin (BSA). Once spheroids formed in this device, they were perfused with the **anti-neoplastic enzyme inhibitor irinotecan (Camptosar[®])** to demonstrate the feasibility of the device for HTS applications. The device allowed cells to grow in 3D and to be perfused at different drug concentrations.

Results

Design of the μ FSCD with a CGG

The dimensions of μ FSCD with a CGG were 4 cm (L) \times 3 cm (W) \times 0.8 cm (H) (Figure 1). The device consists of two layers. The top layer has a 6 mm thickness and contains two inlets (8 mm in diameter), a gradient generator with 150 μ m micro-channels, a culture array with fifty cell injection holes (700 μ m diameters) and five outlets (2 mm diameters). The bottom layer is 2 mm in thickness and contains 50 concave microwells (400 μ m in diameter and 200 μ m in depth), and each channel (C1–C5) has 10 concave microwells (Figure 1B,C).




Figure 1: Design of the μ FSCD with a concentration gradient generator (CGG). View (A) and dimensions (B) of the μ FSCD with a CGG; (C) schematic showing two layers in the μ FSCD with a CGG.

Concentration Gradient on the μ FSCD with a CGG

To quantify the concentration gradient, the same volume (250 μ L) of phosphate-buffered saline (PBS) and 5 μ M fluorescein isothiocyanate (FITC) in PBS were filled into the left and right inlets. Fluorescent images were taken every 8 h and the concentration gradient was maintained for 16 h. Different intensities of fluorescence were observed through the parallel channels (Figure 2A). The FITC intensity of C5 was about 85% of the intensity of 5 μ M FITC and highest among the microwells

(C1–C5). The intensities of the wells decreased gradually from C5 to C1 (Figure 2B).




Figure 2: Concentration gradient of fluorescein isothiocyanate (FITC) in the μ FSCD with a CGG. (A) Fluorescent image of FITC in the channels and concave microwells (C1–C5) 16 h after injection of phosphate-buffered saline (PBS) and FITC into the left and right inlets; (B) Fluorescent intensities of the microwells (C1–C5). $n = 10$.

Spheroid Formation in the μ FSCD with a CGG

Approximately 200 cells were injected through the cell injection holes and approximately 110 cells per microwell were captured. The capturing efficiency was approximately 50%. The captured cells were then cultured in the device. They formed spheroids at D1 and became larger at D2 (Figure 3A). The average diameter of the spheroids ($n = 50$) formed in microwells at D2 was $128.1 \pm 16.6 \mu\text{m}$, thereby indicating that the spheroids were homogeneous (Figure 3B). The distribution of the fifty spheroids as a function of their diameters was also uniform (Figure 3C). Similarly, glioma cell line (U87) were formed homogeneous spheroids and the average diameter of them at D2 was $204.2 \pm 9.7 \mu\text{m}$ (see Supplementary Figure S2).




Figure 3: HCT116 spheroid formation in the μ FSCD with a CGG at different days (0–2). (A) Optical images of spheroids formed in concave microwells. Scale bars, 100 μ m; (B) Spheroid diameters in each channel at D2 ($n = 10$); (C) Spheroid diameter frequency distribution at D2 ($n = 50$).

Drug Screening for Spheroids in the μ FSCD with a CGG

Once spheroids formed, 5 μ M irinotecan which is 8 times higher half maximal inhibitory concentration in 2D culture of HCT116 (Supplementary Figure 1) was injected into the device to generate the concentration gradient of the drug, and the responses of the spheroids in the presence of the gradient were investigated. Figure 4A showed that the spheroids in C1 became larger, while the structures of the spheroids from C2 to C5

became loose, and cells detached from the spheroids. The relative number of HCT116 spheroids in C4 and C5 at D5 was significantly lower than those of the HCT116 spheroids in C1, C2, and C3 (Figure 4B). The roundness of spheroids [26] can be used to determine whether the spheroids were affected by the drug. HCT116 spheroids treated with irinotecan were collapsed. Thus, the roundness of the spheroids in C3, C4, and C5 decreased over time, whereas the roundness of the spheroids in C1 and C2 remained similar (Figure 4C). This indicates that the CGG structure was appropriate to generate the expected drug concentration gradient. These results were supported by live/dead staining images in Figure 4D. Accordingly, C5 yielded the most collapsed spheroid, while the dead cells, which are stained in red, indicated that the concentration of irinotecan in C5 was higher than the concentrations of the other channels. The cell viability was measured by live/dead staining in Figure 4E. At D5, after irinotecan treatment, the cell viability of C5 was the lowest with approximately 63%, while the cell viability in C1 was 98%. The cell viabilities reduced continually following the drug concentration in each channel. These results showed the potential application of the μ FSCD with a CGG in HTS.




Figure 4: Responses of HCT116 spheroids to irinotecan in the μ FSCD with a CGG. (A) Optical images of HCT116 spheroids with the treatment of 5 μ M irinotecan at various days (D2–D5) (Scale bars, 100 μ m). Relative numbers (B) and relative roundness values (C) of HCT116 spheroids in each channel at D2 and D5 ($n = 20$, Student's t -test, * $p < 0.05$, *** $p < 0.001$). Live/dead staining images (D) and cell viability (E) of HCT116 spheroids treated with irinotecan in each channel at D5 (Scale bars, 100 μ m). Calcein AM and EthD-1 were used to stain live and dead cells in green and red, respectively.

Discussion and Conclusions

In this study, we developed a microfluidic spheroid culture device with a CGG for 3D spheroid cultures for high-throughput drug screening. Common microfluidic systems are complicated and use a syringe pump to perfuse nutrients, drugs, and cells [27,28]. Our μ FSCD with a CGG has reservoirs instead of a syringe pump, and could perfuse medium into spheroids in concave microwells. In the μ FSCD with the CGG, HCT116 cells were deposited in concave microwells and were not attached to the PDMS surface because of the BSA coating, and concave shape. Concave shape microwell makes cells aggregate and form spheroid due to the low cell attachment to surfaces [29]. Generally, large spheroids without vascularisation can induce cell death because of oxygen deficiency at the center. Thus, it is important to produce relatively small spheroids. The thickness of the spheroids cultured under optimal nutrient and oxygen conditions ranged from 100 to 220 μ m [30], so the spheroids with diameters in the range of 100 to 150 μ m were selected for drug responses. As a result of the optimisation of the cell density to 5×10^4 cells/mL, relatively similar size spheroids with an average diameter of 120 μ m were obtained in the μ FSCD with a CGG (Figure 2). Although our experiments were performed with the use of a spheroid with an average diameter of 120 μ m in size, the size of the spheroid can be controlled based on the cell density [25]. The uniformness of spheroids is also critical to get precise drug responses. When the spheroids size is not uniform, the permeability of nutrients, gas, and drug is different in each spheroid, and elicits diverse drug responses [31–33]. The μ FSCD with a CGG can generate homogeneous spheroids because it is easy to load a similar number of cells through the cell injection hole.

Gradient generation by CGG provided five different irinotecan concentration conditions and was suited for HTS with a large number of spheroids in parallel. Each generated concentration from CGG can be calculated easily. When only two inlets are included, intermediate values between two neighboring concentrations are generated regardless of what the splitting ratios are and thus the gradient profile is always kept

monotonous [34,35]. Thus, the concentration of treated irinotecan was expected to 0, 1.25, 2.5, 3.75 and 5 μM (C1–C5). CGG has been included in many microfluidic systems for drug screening because it generates various drug concentrations by perfusing a single concentration of the drug. Additionally, its concentration is predictable so it is advantageous to identify the appropriate concentration of the drug [35]. Furthermore, PDMS are proper for extending the duration of the chemical gradient because of their evaporation and liquid absorption functions [36]. In the future, the number of concentration points can be increased to handle more drug concentrations at any time instant.

In addition, the numbers and shapes of spheroids can be easily observed under an optical microscope in label-free conditions owing to the transparent PDMS substrate (Figure 3). Spheroid disaggregation increased after drug treatment. Thus, it was difficult to compare the drug effects by measuring the spheroid diameters. Instead of measuring the diameter, the surface of the spheroids was measured [37], and spheroids which displayed a rough surface were not counted (Figure 3B). The roundness of the spheroids elicited more accurate drug response results as their spheroid shapes were collapsed by the drug (Figure 3C). Furthermore, viability analysis by live/dead staining can provide a sensitive measure of spheroid responses to a particular drug (Figure 4). Spheroids can be easily stained through the cell injection holes.

In conclusion, our μFSCD with a CGG offered a new approach for large-scale drug screening using spheroid microarrays. This approach was based on features of concave microwells connected with a CGG. Given that the μFSCD with the CGG provided a large amount of homogeneous 3D spheroids and several drug concentrations, it can constitute a convenient tool for widespread, parallel processing for the prediction of the effectiveness of the drugs and the determination of the proper cancer drug concentration in patients. Therefore, the proposed μFSCD with a CGG could be useful for clinical samples and could become cost-effective in personalised medicine given that a small number of cells can rapidly form spheroids, and given that the device consumes a small volume of nutrients and drugs.

Materials and Methods

Design and Fabrication of the Device

The μ FSCD with a CGG was designed with AutoCAD (Student Version 2015, Autodesk Korea Limited, Seoul, Korea). The design was printed in a transparent film. For fabricating a mold with the design, the negative photoresist SU-8 (MicroChem Corp., Westborough, MA, USA) was spin-coated on a 4 in silicon wafer for 30 s (top layer-SU-8 2025, 2000 rpm, 40 μ m thickness; bottom layer-SU-8 2150, 2000 rpm, 300 μ m thickness). The coated wafer was exposed to UV through the film using a contact aligner (MDA-400M, MIDAS SYSTEM CO., Daejeon, Korea). The exposed wafer was developed with an SU-8 developer (MicroChem Corp.) to obtain the mold. Then, the mold was treated with trichloro (1H, 1H, 2H, 2H-perfluorooctyl) silane (Sigma-Aldrich, St. Louis, MO, USA) in vacuum conditions at RT for 1 h. Both layers were made from PDMS (Sylgard[®] 184, Dow Corning Corp., Midland, MI, USA) using soft lithography [38]. Fifty cell injection holes (700 μ m diameter), two inlets (8 mm diameter), and five outlets (2 mm diameter) were punched through the top layer. For the manufacturing of the bottom layer with the concave microwells, the cylindrical microwells were filled with PDMS and removed by wiping. Remained PDMS in the cylindrical microwells formed a concave meniscus. Details can be found in a previous publication [39]. Subsequently, both layers were treated with O₂ plasma for 30 s at 90 W, and were then bound to each other.

Demonstration of Concentration Gradients in the μ FSCD with a CGG

The device was treated with 3% (*w/v*) BSA (Sigma-Aldrich) for 1 h at RT and then rinsed three times with PBS (pH = 7.4). The inlets were filled with 250 μ L of PBS, and PBS contained 5 μ M FITC. Fluorescent images were acquired every 8 h using a DeltaVision Elite fluorescence microscope (GE Healthcare, Chicago, IL, USA). Fluorescent intensity profiles over the μ FSCD with a CGG were analysed with ImageJ (Fiji, NIH, Bethesda, MD, USA).

Cells and Their Maintenance

The colon cancer cell line (HCT116) and glioma cell line (U87) was purchased from ATCC (Manassas, VA, USA). HCT116 and U87 was cultured in McCoy's 5A Medium and Minimum Essential Media (Life Technologies, Carlsbad, CA, USA) supplemented with 10% (v/v) fetal bovine serum (HyClone Laboratories, Inc., Logan, UT, USA), 100 units per mL of penicillin (Life Technologies) and 100 µg/mL of streptomycin (Life Technologies). The cells were maintained at 37 °C with 5% CO₂.

Operation of the µFSCD with a CGG

For sanitisation, the µFSCD with a CGG was washed with 70% ethanol through one of the reservoirs, and was rinsed with PBS using a 1 mL syringe. The µFSCD was treated with 3% BSA for 1 h at RT to avoid cell attachment to the bottom. It was rinsed again two times with PBS to remove the uncoated BSA from it. The sanitised device was washed with medium through one of the reservoirs using a 1 mL syringe, and the inlets were filled with 250 µL of medium. Approximately 1 µL of 2×10^4 cells/mL in medium were loaded into the device through each cell injection hole using a pipette. Uncaptured cells in the device were removed by pipetting. The device, which contained the cells, was incubated at 37 °C in a 5% CO₂ incubator for 2 days to generate spheroids. After 2 days, inlets were filled with 250 µL of medium, and contained 5 µM irinotecan and medium. The inlet contents were replaced daily with fresh medium and contained 5 µM irinotecan or medium, and the outlet contents were removed for 3 days.

Live/Dead Staining and Cell Viability Measurements

HCT116 spheroids at D5 after irinotecan treatment were treated with a LIVE/DEAD™ Viability/Cytotoxicity Kit (Molecular Probes, Eugene, OR, USA), as suggested by the manufacturer. The inlets were filled with the mixed live/dead solution and then incubated at 37 °C for 30 min. Green and red fluorescence intensities were measured with ImageJ, and cell viability was

calculated as the ratio of the red with the green fluorescence intensities.

References

1. Fang Y, Eglen RM. Three-dimensional cell cultures in drug discovery and development. *SLAS Discov.* 2017; 22: 456–472.
2. Birgersdotter A, Sandberg R, Ernberg I. Gene expression perturbation in vitro—A growing case for three-dimensional (3D) culture systems. *Semin. Cancer Biol.* 2005; 15: 405–412.
3. Bhadriraju K, Chen CS. Engineering cellular microenvironments to improve cell-based drug testing. *Drug Discov. Today.* 2002; 7: 612–620.
4. Nelson CM, Bissell MJ. Modeling dynamic reciprocity: Engineering three-dimensional culture models of breast architecture, function, and neoplastic transformation. *Semin. Cancer Biol.* 2005; 15: 342–352.
5. Lee J, Cuddihy MJ, Kotov NA. Three-dimensional cell culture matrices: State of the art. *Tissue Eng. Part B Rev.* 2008; 14: 61–86.
6. Baharvand H, Hashemi SM, Ashtiani SK, Farrokhi A. Differentiation of human embryonic stem cells into hepatocytes in 2D and 3D culture systems in vitro. *Int. J. Dev. Biol.* 2004; 50: 645–652.
7. Edmondson R, Broglie JJ, Adcock AF, Yang L. Three-dimensional cell culture systems and their applications in drug discovery and cell-based biosensors. *Assay Drug Dev. Technol.* 2014; 12: 207–218.
8. Montanez-Sauri SI, Beebe DJ, Sung KE. Microscale screening systems for 3D cellular microenvironments: Platforms, advances, and challenges. *Cell Mol. Life Sci.* 2015; 72: 237–249.
9. Ryan SL, Baird AM, Vaz G, Urquhart AJ, Senge M, et al. Drug discovery approaches utilizing three-dimensional cell culture. *Assay Drug Dev. Technol.* 2016; 14: 19–28.
10. Tung YC, Hsiao AY, Allen SG, Torisawa YS, Ho M, et al. High-throughput 3D spheroid culture and drug testing using a 384 hanging drop array. *Analyst.* 2011; 136: 473–478.

11. Madoux F, Tanner A, Vessels M, Willetts L, Hou S, et al. A 1536-well 3D viability assay to assess the cytotoxic effect of drugs on spheroids. *SLAS Discov.* 2017; 22: 516–524.
12. Kessel S, Cribbes S, Dery O, Kuksin D, Sincoff E, et al. High-throughput 3D tumor spheroid screening method for cancer drug discovery using celigo image cytometry. *SLAS Technol.* 2017; 22: 454–465.
13. Zanoni M, Piccinini F, Arienti C, Zamagni A, Santi S, et al. 3D tumor spheroid models for in vitro therapeutic screening: A systematic approach to enhance the biological relevance of data obtained. *Sci. Rep.* 2016; 6: 19103.
14. Friedrich J, Seidel C, Ebner R, Kunz-Schughart LA. Spheroid-based drug screen: Considerations and practical approach. *Nat. Protoc.* 2009; 4: 309–324.
15. Mosaad EO, Chambers KF, Futrega K, Clements JA, Doran MR. The Microwell-mesh: A high-throughput 3D prostate cancer spheroid and drug-testing platform. *Sci. Rep.* 2018; 8: 253.
16. Du G, Fang Q, den Toonder JM. Microfluidics for cell-based high throughput screening platforms—A review. *Anal. Chim. Acta.* 2016; 903: 36–50.
17. Wu G, Doberstein SK. HTS technologies in biopharmaceutical discovery. *Drug Discov. Today.* 2006; 11: 718–724.
18. Chen YC, Lou X, Zhang Z, Ingram P, Yoon E. High-throughput cancer cell sphere formation for characterizing the efficacy of photo dynamic therapy in 3D cell cultures. *Sci. Rep.* 2015; 5: 12175.
19. Chen Y, Gao D, Liu H, Lin S, Jiang Y. Drug cytotoxicity and signaling pathway analysis with three-dimensional tumor spheroids in a microwell-based microfluidic chip for drug screening. *Anal. Chim. Acta.* 2015; 898: 85–92.
20. Anada T, Fukuda J, Sai Y, Suzuki O. An oxygen-permeable spheroid culture system for the prevention of central hypoxia and necrosis of spheroids. *Biomaterials.* 2012; 33: 8430–8441.
21. Liu W, Wang JC, Wang J. Controllable organization and high throughput production of recoverable 3D tumors using pneumatic microfluidics. *Lab Chip.* 2015; 15: 1195–1204.

22. Okuyama T, Yamazoe H, Mochizuki N, Khademhosseini A, Suzuki H, et al. Preparation of arrays of cell spheroids and spheroid-monolayer cocultures within a microfluidic device. *J. Biosci. Bioeng.* 2010; 110: 572–576.
23. Kwapiszewska K, Michalczuk A, Rybka M, Kwapiszewski R, Brzozka Z. A microfluidic-based platform for tumour spheroid culture, monitoring and drug screening. *Lab Chip.* 2014; 14: 2096–2104.
24. Occhetta P, Centola M, Tonarelli B, Redaelli A, Martin I, et al. High-throughput microfluidic platform for 3D cultures of mesenchymal stem cells, towards engineering developmental processes. *Sci. Rep.* 2015; 5: 10288.
25. Fan Y, Nguyen DT, Akay Y, Xu F, Akay M. Engineering a brain cancer chip for high-throughput drug screening. *Sci. Rep.* 2016; 6: 25062.
26. Amaral RLF, Miranda M, Marcato PD, Swiech K. Comparative analysis of 3D bladder tumor spheroids obtained by forced floating and hanging drop methods for drug screening. *Front Physiol.* 2017; 8: 605.
27. Kim C, Bang JH, Kim YE, Lee SH, Kang JY. On-chip anticancer drug test of regular tumor spheroids formed in microwells by a distributive microchannel network. *Lab Chip.* 2012; 12: 4135–4142.
28. Hong JW, Quake SR. Integrated nanoliter systems. *Nat. Biotechnol.* 2003; 21: 1179–1183.
29. Langhans SA. Three-dimensional in vitro cell culture models in drug discovery and drug repositioning. *Front. Pharmacol.* 2018; 9: 6.
30. Sutherland RM. Cell and environment interactions in tumor microregions: The multicell spheroid model. *Science.* 1988; 240: 177–184.
31. Kunz-Schughart L. Multicellular tumor spheroids: Intermediates between monolayer culture and in vivo tumor. *Cell Biol. Int.* 1999; 23: 157–161.
32. Mueller-Klieser W. Multicellular spheroids. *J. Cancer Res. Clin. Oncol.* 1987; 113: 101–122.
33. Hirschhaeuser F, Menne H, Dittfeld C, West J, Mueller-Klieser W, et al. Multicellular tumor spheroids: An underestimated tool is catching up again. *J. Biotechnol.* 2010; 148: 3–15.

34. Dertinger SK, Chiu DT, Jeon NL, Whitesides GM. Generation of gradients having complex shapes using microfluidic networks. *Anal. Chem.* 2001; 73: 1240–1246.
35. Wang X, Liu Z, Pang Y. Concentration gradient generation methods based on microfluidic systems. *RSC Adv.* 2017; 7: 29966–29984.
36. Mulholland T, McAllister M, Patek S, Flint D, Underwood M, et al. Drug screening of biopsy-derived spheroids using a self-generated microfluidic concentration gradient. *Sci. Rep.* 2018; 8: 14672.
37. Lim W, Hoang HH, You D, Han J, Lee JE, et al. Formation of size-controllable tumour spheroids using a microfluidic pillar array (muFPA) device. *Analyst.* 2018; 143: 5841–5848.
38. Xia Y, Whitesides GM. Soft lithography. *Annu. Rev. Mater. Sci.* 1998; 28: 153–184.
39. Lee SA, No DY, Kang E, Ju J, Kim DS, et al. Spheroid-based three-dimensional liver-on-a-chip to investigate hepatocyte-hepatic stellate cell interactions and flow effects. *Lab Chip.* 2013; 13: 3529–3537.

Supplementary Materials




Figure S1: Relative cell viability of HCT116 monolayers with the treatment of irinotecan at different concentrations (0–100 μM) for 72 h. Cell viability was measured using the EZ-cytox Cell Viability Assay Kit (Daeillab Service, Seoul, Korea).




Figure S2: U87 spheroid formation in the μ FSCD with a CGG at different days (0–2). (A) Optical images of spheroids formed in concave microwells. Scale bars, 100 μ m; (B) Spheroid diameters in each channel at D2 (n = 10); (C) Spheroid diameter frequency distribution at day 2 (n = 50).